

Lead per CA AG Protocol for Proposition 65
Quantitative Analysis Report
Inductively Coupled Plasma-Mass Spectrometry

<u>Antacid Tabs</u>	<u>Parts Per Million (ug/g)</u>			<u>Tablet Wt. (g)</u>	<u>Lead (ug/Tab)</u>
	<u>Rep 1</u>	<u>Rep 2</u>	<u>Avg.</u>		
Sample 1	0.027	0.028	0.028	1.8058	0.050
Sample 2	0.029	0.035	0.032	1.8167	0.058
Sample 3	0.031	0.033	0.032	1.7763	0.057
Sample 4	0.029	0.023	0.026	1.8262	0.047

Detection Limit: 0.020

Date Analyzed: 8-11-99

Quality Control Summary

Sample: Antacid Tablets, *****

	<u>Lead Parts Per Million (ug/g)</u>
Sample:	0.268
Duplicate:	0.298
Average:	0.283
Sample RPD:	10.6
Spike Conc:	1.00
MS Result:	1.240
% Recovery	95.7
Control Limits:	
RPD	<25%
% Recovery	80-120%

Date Analyzed: 8-11-99